

Scanner Validation via the SNM Clinical Trials Network Phantom Program

Paul E. Christian
Associate Director - Operations
Molecular Imaging Program
Huntsman Cancer Institute
University of Utah

Chairman, SNM Quality Assurance Committee
Chairman, SNM, CTN Sub-Committee

Phantom Program (CAP, ACNP, SNM)

1973-2010

1973A	Geometric
1973BA	Liver
1974A	Brain
1974B	AP Liver
1975A	AP Thyroid
1975B	Thorax
1976A	Geometric
1976B	Brain
1977A	Liver
1977B	AP Pelvis
1978A	AP Liver
1978B	AP Thyroid
1979A	Liver
1979B	Brain
1980A	Liver
1980B	Brain
1981A	Liver
1981B	Brain
1982A	Thyroid
1982B	Myocardium
1983A	Liver
1983B	Myocardium

1984A	Head and Neck
1984B	Myocardium
1985A	Gallium
1985B	Gated Ejection Fraction
1986A	Geometric ROC Study
1986B	Gated Ejection Fraction
1987A	Liver ROC Study
1987B	Gated Cardiac Arrhythmia
1988A	Feet AP + Lat. ROC Study
1988B	Gated Cardiac Cardiomyop.
1989A	Gated Cardiac Tachycardia
1989B	Renal Dynamic
1990A	Gated Cardiac(akinesis)
1990B	Renal Dynamic
1991A	Pelvis
1991B	Hepatobiliary
1992A	Gated Cardiac Arrhythmia
1992B	Geometric SPECT
1993A	Lung V/P Study
1993B	Brain SPECT Coronal
1994A	Geometric ROC Study
1994B	Cardiac SPECT Quant.

1995A	SPECT QC
1995B	Spinal SPECT/Planar
1996A	Rest/Stress myocardial
1996B	SPECT/Planar Renal
1997A	Skeletal Study
1997B	Medium Energy SPECT
1998A	Cardiac SPECT
1998B	Breast Planar
1999A	Lung Perfusion SPECT
1999B	Dual Isotope Cardiac
2000	Cardiac SPECT
2001	SPECT Lumbar Spine
2002	SPECT Chest
2003	SPECT Cardiac
2004	Gated SPECT Cardiac
2005	SPECT Spine
2006	SPECT Test Phantom
2007	Lung Perfusion
2008	Thyroid
2009	Cardiac TID
2010	Gastric Emptying
2011	PET/CT
2012	SPECT Cardiac

Clinical Simulation Phantoms

Evaluate: Image performance

- Acquisition
- Processing
- Interpretation

Purpose of PET phantom

- Resolution
- Contrast
- Noise
- Quantitative accuracy
- PET/CT alignment
- Attenuation correction






Benefits to VA Medical Centers

- Objective review of imaging
- Global assessment of procedure accuracy
- Evaluate equipment performance
- Evaluate imaging protocols
- Review accuracy of quantitative measurements
- Identify poorly performing equipment
- Identify staff that needs education
- Ensure quality services to patients

Scanner Validation Purpose:

Ensure scanner meets certain performance criterion needed for multi-center clinical trials via a clinical simulation phantom exercise.

Standardization and Harmonization of image quality and quantitative accuracy



Chest Phantom

- Clinical simulator to measure:
 - Lesion detectability
 - Lesion quantitation
- Fill with F-18 FDG
- Precision Filling Technique
- Concentration ($0.14\mu\text{Ci/ml}$)
- 4:1 Lesion/Background Ratio



Imaging Process

- Fill phantom (precise technique)
 - Performing imaging according to prescribed protocol
 - Image after specific time delay
 - Reconstruction parameters
 - MD to identify lesions
 - Submit images for review
-
- Quantitative measurements (SUV_{max} and SUV_{ave})
 - At image quality core lab
 - At site



Demonstration Project -11 scanners

SUV Measurements

<u>Location</u>	<u>Ave SUV</u>	<u>Range SUV</u>	<u>StdDev</u>	<u>COV</u>
Background (1.0)	0.99	0.97 – 1.09	0.04	3.7%
<u>Lesion (4.0)</u>	<u>Max SUV</u>			
20mm	3.00	2.41 – 4.06	0.51	17.0%
15mm	2.58	2.02 – 3.54	0.44	17.0%
A 10mm	1.43	1.08 – 2.03	0.28	19.8%
B 10mm	1.50	1.09 – 2.25	0.30	20.1%
C 10mm	1.60	1.82 – 1.32	0.17	10.7%


CTN Reconstruction Work Group

Standardized/Harmonized image quality and quantitation

Challenges

- Operational:
 - Ship phantom multiple times
 - Sites must image and ship quickly
 - DICOM CD images not readable with commercial software (FDA approved)
- Site Pitfalls:
 - Find time in clinical schedule
 - Made CD on third party system or PACS (44% repeat)
 - Scanner mis-calibrated (absolute activity) (11% repeat)
 - Filled phantom incorrectly (7% repeat)
 - Data entry (Patient weight, injection time, dose) (10%)
 - Dose calibrator error (3%)

New PET/CT Systems





Quality Control

- Dose Calibrator
- CT Scanner
 - CT checks
 - Calibration
 - Inspection
- PET Scanner
 - Blank scan
 - Constancy
 - PET and CT alignment
 - Record of PMs and calibrations



Scanner Constancy



One Year

Courtesy: Paul E. Christian

CTN Scanner Validation



- **Validated scanners** **32**
- **Sites currently imaging phantoms** **19**
- **Sites waiting for phantoms** **>17**

Scanner Validation

Submit Scanner Questionnaire

- Scanner make, model, mode, etc
- Dose calibrator model, etc
- Site contact personnel information

Receive and Image Phantom

Submit Scanning Information to Core Lab

- Submit imaging form
- Submit DICOM images on CD

Core Lab Reviews Images and Data

Site Completion of SUV Measurements

Core Lab Reviews Site SUV Measurements


Site Validation Complete

Site Receives Scanner Validation Certificate

Scanner Validation Sub-Committee

2010 Goals:

Testing brain imaging phantom



Characteristics and Advantages

- Uniformity evaluation
- Resolution pattern of hot and cold rods
- Clinical simulation of central brain region

Scanner Validation Sub-Committee

Testing myocardial imaging phantom

